

## Application Memo

# Measurement of Salt Content in Sweet Sake (Mirin)

Industry	Food
Instrument	Automatic potentiometric titrator
Measurement method	Precipitation titration
Standards	

### 1. Overview

Salt concentration in the diluted sample is measured by titration with 0.1mol/L silver nitrate solution. The endpoint is the maximum inflection on the titration curve. The salt concentration is calculated from the titration volume of the silver nitrate solution.


### 2. Apparatus

Main unit	Automatic potentiometric titrator (preamplifier STD)
Electrode	Silver electrode Mercury sulfate reference electrode

### 3. Reagents

Titrant	0.1 mol/L silver nitrate solution
Solvent	Pure water

### 4. Example



—Measurement results—			
	Sample (mL)	Titration (mL)	Conc. (%)
1	5.0	2.4626	0.2881
2	5.0	2.4724	0.2892
3	5.0	2.4588	0.2876
Average			0.2883
SD			0.0008
RSD(%)			0.3

-Titration curve-

Please feel free to contact us for any further information.

<Contact>Kyoto Electronics Manufacturing Co., Ltd.  
Overseas Sales & Marketing Sect.

<http://www.kyoto-kem.com/en/contact/form.php>